

DENTAL

Osteoplug®

Alveolar Ridge and Socket Preservation

Surgical Guide



Surgical Guide

1

Tooth Extraction

- Atraumatic extraction techniques should be used with attempts to preserve all of the remaining alveolar bone adjacent to the tooth.



2

Surgical Site Preparation

- Thorough curettage of socket.
- Rinse with copious saline or chlorhexidine irrigation for clarity.
- Preparing of local mucogingival flap for the later step of primary wound closure.



Warning: This description is not sufficient for immediate application of the surgical device. Instruction by a surgeon experienced in handling the surgical device is highly recommended.

Surgical Guide *cont'd*

3 Osteoplug® Preparation & Insertion

- To select the proper size of the Osteoplug®, measurement of the socket could be using the periodontal probe to the extracted root fragment. Osteoplug® can be cut/trimmed using surgical scissors or blade according to the desired width and length.
- Osteoplug® should replace the 2/3 top of the extraction socket and tightly fit. For multi-rooted tooth extractions, more than one piece of Osteoplug® may be used to fill the entire sockets.
- The top plane of the Osteoplug® to level with the bone socket. Simple trimming can be done using surgical scissors or blade.
- Allow blood to flow in and create clot inside the scaffold.
- Other biologically active substances such as platelet rich plasma (PRP), bone marrow aspirate (BMA) to enhance bone regeneration can be inserted into the scaffold at this stage.



4 Closure of the surgical site

- Insertion of barrier membrane over the socket.
- Primary wound closure without tissue tension.



Warning: This description is not sufficient for immediate application of the surgical device. Instruction by a surgeon experienced in handling the surgical device is highly recommended.

Disclaimer

Osteopore® does not provide medical advice, diagnosis or treatment. Information contained in this surgical technique guideline have been provided for general information purposes only. This document is not intended to be a comprehensive document on surgical techniques. The information included cannot and should not replace the independent medical judgment of the treating physician. Decisions on appropriate treatment and surgical technique are the treating physician's responsibility and depend on the physician's training, medical knowledge and the case-specific conditions. Successful implantations are technique sensitive. Sound surgical judgments should be used in the selection of all products. The managing physician should inform the patient regarding the potential benefits, risks and complications associated with the treatment.

Although Osteoplug® (hereinafter referred to as "product") has been manufactured under carefully controlled conditions, Osteopore® has no control over the conditions under which the product is used. Osteopore®, therefore, disclaims all warranties, both expressed and implied, with respect to the product, including, but not limited to, any implied warranty of merchantability or fitness for a particular purpose.

Osteopore® shall not be liable to any person or entity for any medical expenses or any direct, indirect, incidental or consequential damages caused by any use, defect, failure or malfunction of the product, whether a claim for such damages is based upon warranty, contract, tort or otherwise. No person has any authority to bind Osteopore® to any representation or warranty with respect to the product. The exclusions and limitations set out above are not intended to, and should not be construed so as to contravene mandatory provisions of applicable law.

If any part or term of this Disclaimer of Warranty is held to be illegal, unenforceable or in conflict with applicable law by court of competent jurisdiction, the validity of the remaining portions of this Disclaimer of Warranty shall not be affected, and all rights and obligations shall be construed and enforced as if this Disclaimer of Warranty did not contain the particular part or term held to be invalid.

For Professional Use.

CAUTION: See instructions for use for full prescribing information, including indications, contraindications, warnings, and precautions. Osteopore® devices placed on the European market meeting the essential requirements referred to in Article 3 of the Medical Device Directive 93/42/EEC bear the CE marking of conformity.

Osteopore®, Osteopore International, the stylized Osteopore® logo, are trademarks of Osteopore International Pte Ltd.

© 2020 Osteopore®. All rights reserved.

Osteopore®

Empowering Natural Tissue Regeneration

Osteopore International | SINGAPORE

2 Tukang Innovation Grove #09-06, JTC MedTech Hub, Singapore 618305

MCN 04 013B XE Rev 0